

Hospital Bed Electric System

M-07.3 EL



SPECIFICATION :

► **Main Frame**

Made of square steel pipe $\varnothing 30 \times 60 \times 1,2$ mm,
steel pipe $\varnothing 32$ mm $\times 1,2$ mm
and steel plat 1,2 mm thickness

► **Back Raise and Knee Raise Frame**

Made of steel pipe $\varnothing 1\frac{1}{4}$ ",
square steel pipe $\varnothing 25 \times 25$ mm welding proses.

► **Panel (Head and Foot panel)**

Removable head and foot panel
made of plastic ABS

► **Mattress Base**

Made of plastic ABS

► **Adjusment Range**

Back Raise, Knee Raise and Hi-Low, adjustable
by electric system. (Hi-Low used gas spring)

► **I.V Stand (2 or 4 Hooks)**

Made of stainless steel pipe

► **Side Rail**

Made of plastic ABS

► **Castor**

4 castors 5" from rubber
(two castors with brake)

► **Finishing**

Powder Coating

► **Max Load**

± 200 kg

► **Dimension**

210(L) \times 92,5(W) \times 100(H) cm